


CERTIFICATE OF ANALYSIS
T2 Toxin in Acetonitrile LCMS grade

This document is designed, and the certified values and uncertainty are determined in accordance with ISO Guide 31, ISO Guide 34, ISO Guide 35, AOAC, and Eurachem/ CITAC Guides.

Description of the Reference Material (RM)

	Product name:	T2 Toxin		
	Product number:	FIA000299		
	CAS number:	T2 Toxin	21259-20-1	
	Lot number:	T218042301		
	Expiry date:	23-Apr-2026		
	Certified value (s):	T2 Toxin	100,32 ± 4,23	µg/mL
	Physical description:	Clear solution of toxin in Acetonitrile LCMS grade		
	Packing	Amber glass vial filled with 1 mL of solution		
	Storage conditions	≤ -10°C		
	Matrix and starting material:	This material was prepared with/from:		
		Acetonitrile LCMS Grade	Batch:	0001204102BS
		T2 Toxin	Internal ID:	T008

Intended use of the RM:

For laboratory use for R&D purposes only. The main purpose of this material is for analytical instrument calibration (e. g. external calibration, standard addition). Not for drug, household or other uses.

Instruction for the correct use of the RM:

The vial should be stored in a dark place at ≤ -10°C . Before usage of the RM, allow the vial to warm to room temperature. The expiry date of this RM is based on the current knowledge and holds only for proper storage conditions in the originally closed vials / packages. Solutions prepared for calibration purpose should be protected from exposure to light. Discard solutions after use in accordance with appropriate safety regulations for chemical substances.

Hazardous situation:

H225 : Flammable liquid - Category 2 - Highly flammable liquid and vapour
 H301 : Acute toxicity - Oral - Category 3 - Toxic if swallowed
 H311 : Acute toxicity - Dermal - Category 3 - Toxic in contact with skin
 H319 : Eye irritation - Category 2 - Causes serious eye irritation
 H331 : Acute toxicity - Inhalation - Category 3 - Toxic if inhaled

In case of accident or if you feel unwell, seek medical advice immediately (show the label where possible). Avoid exposure. Wear suitable protective clothing.

Safety measures:

Special care must be taken when manipulating this standard. Avoid contact with eyes, skin and clothing. Avoid prolonged or repeated exposure. Use only in a chemical fume hood. Safety shower and eye bath must be near. In case of spills, cover and absorb with an inert dry material such as dry-lime, sand or soda ash and place in an appropriate waste disposal container.

Keep container tightly closed. Do not store in direct sunlight. Keep away from heat, sparks, flame and incompatible material. Storage area should be cool, dry and away from incompatible materials.

Further information:

Further information is available in the SDS available online (downloading page: www.fianovis.com/documentation (documentation section)). Final users should conduct their own investigations to determine the suitability of the information for their particular research purposes. Under no circumstances will the supplier of this RM be held responsible for any damage resulting from handling or contact with the product.

Traceability

The certified values are based on the results of analytical techniques previously used for purity assessment of solid mycotoxins. High purity material represents a practical realization of concentration units, through conversion of mass to molar quantity.

Calculation of certified values and associated uncertainties

This calibrant is certified on weighing procedure. Toxin is weighed then dissolved in Acetonitrile LCMS grade. Mass concentration calculation is based on weigh, purity and dilution step.

The pipet was calibrated with traceability to national and international standards (Dakks & ilac-MRA). All weights used for metrological control are connected to national and international standards. The weights are calibrated by an accredited laboratory.

$$C (\mu\text{g/mL}) = \frac{m \times P}{V}$$

Toxin	Source				Standard uncertainty
T2 Toxin	Purity		86,41	%	1,68
	Powder	weighing	5805,00	μg	0,10
	Liquid solution C _{ss}	concentration	1116,49	μg/mL	8,97
	Dilution	volume	50,71	mL	0,06
$Combined_u = \sqrt{\left(\frac{u_p}{P}\right)^2 + \left(\frac{u_m}{m}\right)^2 + \left(\frac{u_D}{D}\right)^2 + \left(\frac{u_C}{C}\right)^2}$					0,02
$Concentration_{Toxin} = \frac{Toxin\ mass}{V_{D1}}$ μg/mL					100,32
Total expanded uncertainty (using a coverage factor k=2)					4,23

Notes: The purity of the mycotoxin used for this RM was determined by liquid chromatography. Following the Guide to the Expression of Uncertainty in measurement (GUM) the expanded uncertainty of toxin level is obtained by multiplication with a coverage factor K for which 2 is usually chosen to obtain a confidence level of 95 %.

Quality control


Confirmation of the certified concentration by HPLC-PDA

The certified concentrations of the prepared solution was confirmed by HPLC-PDA against a reference batch.

Chromatogram

Chromatographic conditions:

Column:	Luna C18 150 x 4,6 mm 5 μm		
Mobile phase:	H2O + 0.1% acide acetique / ACN / Isocratic : 50%B / 50%C		
Flow mL/min:	1,50		
Temperature °C:	30,00		
Detector	PDA		
T2 Toxin	102,25	± 1,20	μg/mL
Mean of 6 replicates measurement against reference batch, confidence interval with P = 95%			



Chromatogram of Toxins

References:

- a-ISO GUIDE 31:2015, Reference Materials - Contents of certificates, labels and accompanying documentation.
- b-ISO GUIDE 34:2009, General requirements for the competence of reference material producers
- c-ISO GUIDE 35:2006, Reference materials - General and Statistical Principles.
- d-ISO/IEC Guide 98-3:2008 Uncertainty of measurement-Part 3 : Guide to the expression of uncertainty in measurement (GUM:1995)
- e-Eurachem/CITAC guide (2019), Traceability in Chemical Measurement.
- f-Eurachem/CITAC guide (2012), Quantifying Uncertainty in Analytical Measurement.
- g-AOAC Official Method 970.44-1971 - Preparation of Standards for Mycotoxins.

This document was computer generated and is valid without signature

Prepared by: 26-Apr-2024
Quality Control

Date: CLERMONT Alexandre